

Wt. sets
used:

CALEY & WHITMORE CORPORATION

"LABORATORY INSTRUMENT SERVICE SPECIALISTS"

18 HIGHLAND AVENUE • SOMERVILLE, MA 02143

(617) 623-7430

DUE
6-30-13

SERVICE SUMMARY LISTING SHEET

FIRM INSTITUTION

State Health Labs

NUMBER

ADDRESS

DATE

6-6-12

P.O. NUMBER

DEPT.

Amherst

MAKE
CERTS
TO: 7

PERSON

413-545-2607
Jim Hanchette

S.

ITEM	MANUFACTURER	TYPE	MODEL#	SERIAL #	ID #	L.N.	WTS.	COMMENTS	PART INSTAL	
1	Sartorius	T	L2200S	35110064		$\pm 10\text{mg}$	PRE: 2000.01g POST: 2000.00g	1000.01g 1000.00g	500.00g 500.00g	(210)
2	Ohaus	A	E12140	0063118063041		$\pm 0.1\text{mg}$	PRE: 99.9974g POST: 100.0000g	49.9991g 50.0000g	9.9999g 10.0000g	(210)
3	Ohaus	A	E12140	0063118063051		$\pm 0.1\text{mg}$	PRE: 100.0004g POST: 100.0000g	50.0004g 50.0000g	10.0003g 10.0000g	(210)
4	Ohaus	A	E12140	K1031127133244P		$\pm 0.1\text{mg}$	PRE: 99.9990g POST: 100.0000g	49.9998g 50.0000g	10.0000g 10.0000g	(210)
5	AND	T	FX2000	5228743		$\pm 10\text{mg}$	PRE: 2000.09g POST: 2000.00g	1000.05g 1000.00g	500.03g 500.00g	(210)
6	Ohaus	A	E12140	0063118063044		$\pm 0.1\text{mg}$	PRE: 99.9942g POST: 100.0000g	49.9970g 50.0000g	9.9996g 10.0000g	(210)
7	AND	T	FX2000	5215598		$\pm 10\text{mg}$	PRE: 1999.96g POST: 2000.00g	999.95g 1000.00g	500.00g 500.00g	(210)
8	Mettler	T	PE16	D60658		$\pm 0.1\text{g}$	PRE: 5000.0g POST: 5000.0g	3000.0g 3000.0g	1000.0g 1000.0g	(16K)
9	AND	A	FR300	6200587		$\pm 0.1\text{mg}$	PRE: 99.9933g POST: 100.0000g	49.9966g 50.0000g	9.9991g 10.0000g	(210)
10	AND	T	FX2000	5210160		$\pm 10\text{mg}$	PRE: 2000.06g POST: 2000.00g	1000.03g 1000.00g	500.01g 500.00g	(210)

CUSTOMER
SIGNATURE:

SERVICE
ENGINEER:

CB